

SINGLE FREQUENCY LASER External Cavity Diode Laser

General Product Information

Product	Application
770 nm mini-ECL	Spectroscopy (K D1 line)
with hermetic 14-Pin Butterfly Housing (RoHS compliant)	Metrology
including Monitor Diode, Thermoelectric Cooler and Thermistor	Quantum Technology
with integrated Beam Collimation	



Absolute Maximum Ratings

Parameter	Symbol	Unit	min	typ	max
Storage Temperature	Ts		-40		85
Operational Temperature at Case	T _C	°C	-40		85
Operational Temperature at Chip	T_{chip}	°C	-5		50
Forward Current	١ _F	mA			200
Reverse Voltage	V _R	V			2
Output Power	P _{opt}	mW			100
TEC Current	I _{TEC}	А			1.1
TEC Voltage	V _{TEC}	V			2.8

Recommended Operational Conditions

Parameter	Symbol	Unit	min	typ	max
Operational Temperature at Case	T _{case}	°C	-20		65
Operational Temperature at Chip	T _{chip}	°C	0		40
Forward Current	I _F	mA			180
Output Power	P _{opt}	mW	20		80

Characteristics

= 25° C at BOL

Parameter	Symbol	Unit	min	typ	max
Center Wavelength	λ _C	nm	769	770	771
Target Wavelength	λ_{T}	nm		770.1	
Linewidth	Δλ	MHz		0.1	0.3
Mode-hop free Tuning Range	$\Delta\lambda_{tune}$	pm		15	
Output Power	Popt	mW			80
Sidemode Suppression Ratio	SMSR	dB	30	50	
Temp. Coefficient of Wavelength	$d\lambda/dT$	nm/K		0.008	
Current Coefficient of Wavelength	dλ / dl	nm/mA		0.001	

Measurement Conditions / Comments Stress in excess of one of the Absolute Maximum Ratings may damage the laser. Please note that a damaging optical power level may occur although the maximum current is not reached. These are stress ratings only, and functional operation at these or any other conditions beyond those indicated under Recommended Operational Conditions is not implied.

Measurement Conditions / Comments	
measured by integrated Thermistor	

Measurement Conditions / Comments

reached within TLD = 0° C \dots 45° C
measured in the time scale of 1 ms
By current tuning, at target wavelength
Popt = 80 mW


© eagleyard Photonics GmbH Rudower Chaussee 29 (IGZ) This data sheet is subject to change without notice.

D-12489 Berlin GERMANY fon 49.30.6392.4520 fax 49.30.6392.4529 www.toptica-eagleyard.com info@toptica-eagleyard.com

2023-10-25



= 25° C at BOL Characteristics Parameter Symbol Unit min typ max Laser Current \mathbf{I}_{LD} mΑ 180 Slope Efficiency 0.8 mW/mA 0.6 1 η Threshold Current I_{th} 90 mΑ $\Theta_{||}$ 2 Divergence parallel mrad 2 Divergence perpendicular Θ_{\perp} mrad Beam Diameter horizontal d|| mm 1 1.2 ${\rm d}_{\! \perp}$ 0.8 Beam Diameter vertical 1.2 mm Degree of Polarization DOP % 90



Monitor Diode

Parameter	Symbol Unit	min	typ	max
Monitor Detector Responsivity	I _{mon} / Ρ _{opt} μΑ/mV	V 0.5		100

Measurement Conditions / Comments 5 V

Thermoelectric Cooler

Parameter	Symbol	Unit	min	typ	max
Current	I _{TEC}	А		0.4	
Voltage	U _{TEC}	V		1.3	
Power Dissipation (total loss at case)	Ploss	W		0.5	
Temperature Difference	ΔT	К			50

Thermistor (Standard NTC Type)

Parameter	Symbol	Unit	min	typ	max
Resistance	R	kΩ		10	
Beta Coefficient	β			3892	
Steinhart & Hart Coefficient A	А		1.	1293 x 10 ⁻	3
Steinhart & Hart Coefficient B	В		2.	3410 x 10 ⁻	-4
Steinhart & Hart Coefficient C	С		8.	7755 x 10 ⁻	-8

Measurement Conditions / Comments Popt = 80 mW, ΔT = 20 K Popt = 80 mW, ΔT = 20 K Popt = 80 mW, ΔT = 20 K Popt = 80 mW, ΔT = |Tcase - TLD|

Aeasurement Conditions / Comments	_
$R_1/R_2 = e^{\beta}(1/T_1 - 1/T_2)$ at Tchip = 0° 50°	° C
$/T = A + B(ln R) + C(ln R)^{3}$	
: Temperature in Kelvin	
R: resistance at T in Ω	

fon 49.30.6392.4520 www.toptica-eagleyard.com fax 49.30.6392.4529 info@toptica-eagleyard.com

© eagleyard Photonics GmbH Rudower Chaussee 29 (IGZ) This data sheet is subject to change without notice. D-12489 Berlin GERMANY

ΤΟΡΤΙCΑ







SINGLE FREQUENCY LASER External Cavity Diode Laser


Pin Assignment

hermoelectric Cooler (-)
ase
ot connected
aser Diode Cathode
aser Diode Anode
ot connected
ot connected
2



Package Drawings





interemision


AIZ-20-1029-0928





SINGLE FREQUENCY LASER **External Cavity Diode Laser**

Typical Measurement Results



Unpacking, Installation and Laser Safety

Unpacking the laser diodes should only be done at electrostatic safe workstations (EPA). Though protection against electro static discharge (ESD) is implemented in the laser package, charges may occur at surfaces. Please store this product in its original package at a dry, clean place until final use. During device installation, ESD protection has to be maintained.

A laser diode is sensitive against optical feedback, so an optical isolator may be required in order to avoid any disturbance of the emission spectrum. Operating at moderate temperatures on proper heat sinks will contribute to a long lifetime of the diode.

Avoid direct and/or indirect exposure to the free running beam. Collimating and focussing the free running beam with optics as common in optical instruments will increase threat to the human eye.

Each laser diode will come with an individual test protocol verifying the parameters given in this document.

Performance figures, data and any illustrative material provided in this specification are typical and must be specifically confirmed in writing by eagleyard Photonics before they become applicable to any particular order or contract. In accordance with the eagleyard Photonics policy of continuous improvement specifications may change without notice.









Complies with 21 CER 1040 10 and 1040 40



